

TECHNICAL DATA SHEET

Tristel Solo For Ultrasound



DESCRIPTION

Tristel Solo for Ultrasound disinfectant wipes are designed specifically for the simple and effective disinfection of skin surface probes and ultrasound station monitors, keyboards, probe holders and cables. These disinfectant wipes are based on a Quaternary Ammonium Compound (QAC) in the form of didecyl dimethyl ammonium chloride. To aid cleaning and soil removal, the formulation incorporates two non-ionic surfactants and one chelating agent.



TECHNICAL SPECIFICATION

Quantity	Pack of 100 wipes Case of 12 x 100 wipes
Wipe Size	27 x 20cm
Wipe Thickness	40gsm
Wipe Material	Polycellulose
Test Standard	CE certified
Recommended Use	Skin Surface Probes Monitors Keyboards Probe Holders Cables
Material Compatibility	Rubber Compounds Plastics Mild Steel Stainless Steel Aluminium Copper Brass

FEATURES

- QAC-based disinfectant
- Effective against a wide range of micro-organisms
- For non-invasive ultrasound probes and equipment
- High quality wipes
- 40gsm Polycellulose material
- CE certified

TECHNICAL DATA SHEET



Tristel Solo For Ultrasound

Standards and Directives:

Tristel solo for Ultrasound is CE marked as a Class IIb Medical Device in accordance with the European Medical Devices Directive 93/42/EEC and the 2007/47/EC amendments

Efficacy:

Tristel solo for Ultrasound has demonstrated efficacy against the micro-organisms listed below:

- Enveloped viruses
- Staphylococcus aureus
- Pseudomonas aeruginosa
- Candida albicans
- Methicillin-resistant Staphylococcus aureus (MRSA)
- Vancomycin-resistant Enterococci (VRE)
- Carbapenem-resistant Enterobacteriaceae (CRE)
- Multi-drug resistant Acinetobacter baumannii (MDRAB)
- Escherichia coli
- Enterococcus hirae